



Certificate of Analysis

Secondary Certified Reference Standard NIST traceable

Zinc (Zn) Standard Solution 1000 mg/L (in 5% HNO₃)

Molecular Formula:	NA	Lot No.:	Z-001
Molecular Weight:	NA	Cat No.:	NIST -014
Packing :	100 mL	Dt. of Issue of Certificate:	18/12/2025
CAS No.:	NA	Dt. of Mfg:	Dec 2025
Traceability:	To NIST SRM 3168a	Exp. Dt.:	Nov 2027

TABLE :1 Certified Value

Measurand / Constituent	Certified Value	Expanded Uncertainty (K=2)
Zinc mass fraction (mg/kg)	1034	± 7
Zinc Concentration (mg/L)	1000	± 7

- This Zinc Standard Solution is prepared by using ISO 17034:2016 certified RM & comparing with Zinc Standard solution of NIST SRM 3168a.
- The Zinc Standard Solution is prepared gravimetrically using ISO 17034:2016 CRM Zinc metal as starting material, whose purity is determined by primary method. The Zinc Standard Solution is prepared in 0.8 mol/L nitric acid. One unit of this Reference Material consists of 100 ml Zinc Standard Solution packed in tightly sealed white HDPE bottles. Certified value of Zinc Standard Solution along with the associated uncertainty was calculated at 95% confidence level with coverage factor $k=2$, considering major sources of uncertainty including measurement replication, instrument background, buoyancy correction.

Traceability: The test result is traceable to Standard Reference Material of National Institute of Standards and Technology (USA), SRM 3168a.

Intended use: This reference material is intended to be used as a standard for AAS/ICP-OES & ICP-MS quantitative determination of Zinc and calibration of instruments.

Precautions:

- The sampling should be done using clean and dry pipette/micropipette.
- Excess sample must be rejected and must not be put back into the original bottle.
- Gloves, masks, goggles and other general precautionary measure should be used while handling the reference material.

Accreditation: Sabiha Research And Development ltd. is accredited to ISO 17025:2017.

Expiry: The validity of the standard is 2 years from date of manufacturing, However once the bottle is opened the expiry shall be one year from date of opening but with in overall shelf life of the standard.

Do not freeze, store the bottle tightly closed between 5 - 30°C at room temperature protected from light, keep container tightly closed. Do not put back unused solution into original bottle. For laboratory and industrial use only, not for Human/Animal consumption.

QC Supervisor
Sign _____