



Certificate of Analysis

Secondary Reference Standard NIST traceable

Conductivity Solution ~ 20.0 $\mu\text{S}/\text{cm}$ @ 25° C

Molecular Formula:	NA	Lot No.:	C-014
Molecular Weight:	NA	Cat No.:	NIST -042
Packing :	500 mL	Dt. of Issue of Certificate:	20/04/2026
CAS No.:	NA	Dt. of Mfr:	April 2026
Traceability:	To NIST SRM 999c	Exp. Dt.(Refer Note below):	Mar 2028

TABLE :1 Certified Value

Measurand / Constituent	Certified Value	Expanded Uncertainty (K=2)
Conductivity (at 25 °C)	20.0 ($\mu\text{S}/\text{cm}$)	± 1 ($\mu\text{S}/\text{cm}$)

- This conductivity Standard Solution is prepared by using ISO 17034:2016 certified RM & comparing with Conductivity solution prepared from NIST Standard.

Metrological Traceability: The test result is traceable to Standard Reference Material of National Institute of Standards and Technology (USA), SRM 999c potassium chloride.

Conductivity Substance: Potassium Chloride

TABLE :2 Temperature dependence information

Temperature (° C)	Conductivity ($\mu\text{S}/\text{cm}$)	Temperature (° C)	Conductivity ($\mu\text{S}/\text{cm}$)
5	15.62	30	21.80
10	17.13	35	24.10
15	17.77	40	25.70
20	18.99	50	30.00

Accreditation: Sabiha Research And Development Ltd. is accredited to ISO 17025:2017 by the National Accreditation board and testing calibration laboratory under scope for the test method, TPPHB, used to generate the above result. This accreditation deems Sabiha Research competent on a quality systems level and technical level to perform the tests on the scope of accreditation. Sabiha Research has the quality management system in place to ensure that each individual test result generated using TPPHB is technically valid and is supported by appropriate uncertainty measurements.

Intended use: This reference material is intended for use in aqueous solution as a calibration standard for the determination of the conductivity sample.

- Do not freeze, store at room temperature, keep container tightly closed. Do not put back unused solution into original bottle. For laboratory and industrial use only not for Human/Animal consumption

NOTE: Expiry Shall not exceed 12 months from date of opening within the shelf life period.

QC Supervisor

Sign _____